

平成26年度 第1回 生乳検査外部精度管理調査(成分) R管理

番号	R管理	試料1					試料2					試料3				
		脂肪	タンパク	乳糖	TMS	SNF	脂肪	タンパク	乳糖	TMS	SNF	脂肪	タンパク	乳糖	TMS	SNF
	R平均	0.010	0.008	0.010	0.019	0.016	0.010	0.007	0.008	0.014	0.013	0.009	0.006	0.007	0.015	0.012
管理限界	0.026	0.020	0.027	0.049	0.042	0.025	0.019	0.021	0.037	0.033	0.023	0.015	0.019	0.038	0.032	
61	FT6000	0.000	0.000	0.010	0.010	0.010	0.000	0.000	0.000	0.010	0.010	0.000	0.000	0.000	0.010	0.010
62	FT1	0.010	0.010	0.000	0.020	0.000	0.020	0.000	0.010	0.030	0.030	0.030	0.000	0.010	0.040	0.010
63	FT120	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.010	0.010	0.010	0.000	0.000	0.000	0.000	0.000
64	FT120	0.040	—	—	0.040	0.030	0.050	—	—	0.040	0.010	0.040	—	—	0.040	0.000
65	FT120	0.020	0.010	0.010	0.040	0.020	0.010	0.020	0.020	0.070	0.040	0.010	0.000	0.010	0.010	0.010
66	FT2	0.000	0.010	0.000	0.000	0.020	0.000	0.000	0.000	0.010	0.010	0.000	0.000	0.000	0.000	0.010
67	Minor	0.000	0.010	0.010	0.030	0.040	0.010	0.050	0.010	0.050	0.050	0.000	0.010	0.000	0.010	0.000
68	FT120	0.010	0.000	0.020	0.010	0.030	0.010	0.000	0.000	0.010	0.010	0.010	0.000	0.010	0.010	0.020
69	FT120	0.020	—	—	0.010	0.010	0.030	—	—	0.020	0.010	0.010	—	—	0.010	0.000
70	FT120	0.010	0.000	0.010	0.000	0.000	0.020	0.020	0.000	0.030	0.020	0.010	0.000	0.000	0.010	0.010
71	S50	0.010	0.000	—	0.010	0.010	0.010	0.010	—	0.020	0.010	0.010	0.000	—	0.010	0.000
72	133	0.010	0.010	0.010	0.010	0.000	0.000	0.010	0.010	0.010	0.010	0.010	0.010	0.010	0.020	0.010
73	FT120	0.010	0.010	0.010	0.000	0.010	0.000	0.010	0.010	0.000	0.010	0.000	0.000	0.000	0.000	0.010
74	133	0.010	0.010	0.010	0.020	0.020	0.030	0.000	0.000	0.030	0.000	0.030	0.000	0.010	0.030	0.000
75	S50	0.000	0.010	0.000	0.020	0.020	0.000	0.000	0.010	0.000	0.000	0.010	0.000	0.010	0.000	0.010
76	G・SS	0.050	—	—	0.030	0.030	0.050	—	—	0.010	0.060	0.000	—	—	0.020	0.020
77	FT120	0.000	0.000	0.010	0.010	0.020	0.000	0.000	0.000	0.010	0.010	0.000	0.010	0.000	0.020	0.010
78	FT120	0.010	0.000	—	0.000	0.010	0.010	0.000	—	0.000	0.010	0.010	0.010	—	0.000	0.010
79	FT120	0.010	—	—	0.010	0.010	0.010	—	—	0.000	0.010	0.010	—	—	0.010	0.010
80	FT120	0.000	0.020	0.010	0.020	0.020	0.010	0.010	0.000	0.030	0.020	0.010	0.010	0.010	0.020	0.020
81	FT120	0.000	0.000	0.020	0.010	0.010	0.010	0.000	0.020	0.010	0.000	0.000	0.000	0.010	0.010	0.010
82	S50	0.010	0.010	—	0.020	0.020	0.030	0.000	—	0.020	0.010	0.020	0.010	—	0.010	0.010
83	FT2	0.010	0.000	0.010	0.010	0.010	0.020	0.000	0.010	0.010	0.020	0.010	0.010	0.010	0.010	0.020
84	S50	0.010	0.010	0.010	0.020	0.010	0.030	0.010	0.010	0.010	0.020	0.020	0.010	0.010	0.010	0.010
85	FT120	0.020	0.010	0.000	0.020	0.000	0.000	0.000	0.000	0.000	0.000	0.010	0.010	0.010	0.010	0.010
86	ラクトスコープFTIR	0.010	0.010	—	0.010	0.010	0.010	0.010	—	0.010	0.000	0.000	0.010	—	0.010	0.010
87	S50	0.000	0.010	0.010	0.010	0.010	0.000	0.010	0.000	0.010	0.010	0.000	0.000	0.000	0.010	0.010
88	S50	0.010	0.010	0.010	0.020	0.010	0.010	0.000	0.010	0.020	0.010	0.010	0.010	0.000	0.020	0.010
89	FT120	0.000	0.010	0.010	0.010	0.010	0.000	0.000	0.000	0.000	0.010	0.010	0.000	0.000	0.010	0.010
90	FT120	0.010	0.010	0.040	0.040	0.050	0.010	0.050	0.030	0.030	0.040	0.010	0.010	0.030	0.020	0.030
91	ラクトスコープ	0.020	0.010	—	0.030	0.030	0.020	0.010	—	0.010	0.010	0.030	0.010	—	0.020	0.030
92	FT120	0.000	0.010	0.000	0.010	0.010	0.000	0.000	0.010	0.010	0.010	0.000	0.010	0.010	0.010	0.010
93	G・SS	0.000	—	—	0.010	0.010	0.000	—	—	0.030	0.030	0.000	—	—	0.010	0.010
94	S50	0.010	0.010	0.010	0.010	0.020	0.000	0.000	0.010	0.010	0.010	0.000	0.010	0.010	0.020	0.020
95	S50	0.010	0.000	—	0.010	0.010	0.020	0.000	—	0.010	0.010	0.010	0.000	—	0.010	0.000
96	FT2	0.010	0.010	0.000	0.010	0.010	0.010	0.000	0.010	0.010	0.020	0.000	0.010	0.010	0.000	0.010
97	G・SS	0.050	—	—	0.050	0.090	0.050	—	—	0.050	0.020	0.000	—	—	0.050	0.050
98	G・TMSC	0.050	—	—	0.060	0.020	0.020	—	—	0.050	0.030	0.000	—	—	0.060	0.060

記号説明

- G ゲルベル法
- SS スマートシステム
- TMSC TMSチェッカー