

平成26年度 第3回 生乳検査外部精度管理調査(成分) Z-スコア

| 番号 | Z-スコア | 試料1 | | | | | 試料2 | | | | | 試料3 | | | | |
|-----|-------------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|
| | | 脂肪 | タンパク | 乳糖 | TMS | SNF | 脂肪 | タンパク | 乳糖 | TMS | SNF | 脂肪 | タンパク | 乳糖 | TMS | SNF |
| | | 標準値(%) | 3.84 | 3.47 | 4.49 | 12.78 | 8.95 | 3.95 | 3.35 | 4.49 | 12.78 | 8.83 | 3.82 | 3.28 | 4.52 | 12.61 |
| | 標準偏差 | 0.023 | 0.018 | 0.022 | 0.037 | 0.031 | 0.025 | 0.016 | 0.018 | 0.036 | 0.026 | 0.020 | 0.020 | 0.022 | 0.034 | 0.030 |
| 61 | FT6000 | -0.870 | -0.556 | 0.000 | -1.081 | -0.968 | -0.800 | -0.625 | 0.000 | -1.111 | -0.769 | -0.500 | -0.500 | 0.000 | -0.882 | -0.667 |
| 62 | FT1 | -2.174 | 0.000 | 0.000 | -1.622 | -2.581 | -2.000 | 0.000 | 0.000 | -1.667 | -1.923 | -2.000 | 0.000 | -1.364 | -0.588 | -0.667 |
| 63 | FT120 | -0.435 | 1.667 | -0.909 | 1.081 | 0.968 | 0.800 | 1.250 | 0.000 | 1.667 | 1.538 | 1.000 | -1.000 | -0.455 | 0.294 | -0.333 |
| 64 | FT120 | -1.739 | — | — | 0.000 | 0.968 | 0.800 | — | — | 0.556 | 0.000 | -1.000 | — | — | -0.882 | -0.333 |
| 65 | FT120 | -1.304 | 0.556 | -1.364 | -0.811 | -0.323 | -0.800 | 0.000 | -1.111 | -0.833 | -0.385 | -1.500 | 0.000 | -0.909 | -1.176 | -0.333 |
| 66 | FT2 | -0.435 | 0.556 | 0.909 | 0.000 | 0.323 | -0.800 | 0.000 | 0.556 | -0.556 | 0.769 | -1.000 | 0.000 | 0.455 | -0.588 | 0.667 |
| 67 | Minor | 2.609 | 2.778 | -1.364 | 3.243 | 1.290 | 2.400 | ※ | -1.111 | 3.056 | 2.692 | 0.500 | 3.500 | -0.909 | 2.647 | 2.333 |
| 68 | FT120 | 0.000 | 1.111 | -0.455 | -0.541 | -0.968 | 0.800 | 0.625 | -0.556 | -0.556 | -0.385 | 0.500 | 0.500 | -0.455 | -0.588 | 0.000 |
| 69 | FT120 | 0.435 | — | — | 0.541 | 0.000 | 0.000 | — | — | 0.556 | 0.769 | 1.500 | — | — | 0.294 | -0.667 |
| 70 | FT120 | 0.435 | 0.000 | -2.273 | -0.811 | -1.290 | -0.400 | -1.250 | ※ | -1.944 | -2.308 | 1.500 | -2.500 | ※ | -1.765 | -2.667 |
| 71 | FT2 | -0.870 | 0.000 | 0.455 | -0.270 | -1.290 | -1.200 | -0.625 | -0.556 | -0.278 | -1.154 | -1.000 | -1.000 | -0.909 | 0.000 | -1.667 |
| 72 | 133 | 0.000 | -0.556 | 0.455 | 0.541 | 0.323 | 0.400 | 0.000 | 0.000 | 0.278 | 0.385 | 0.000 | 0.000 | -0.455 | 0.294 | 0.333 |
| 73 | FT120 | 0.000 | -1.111 | -0.909 | -0.811 | -1.613 | -1.600 | -0.625 | -0.556 | -1.667 | -0.769 | -0.500 | -1.000 | -0.909 | -1.471 | -1.333 |
| 74 | 133 | -1.304 | 0.000 | 0.455 | -0.270 | 0.645 | -0.800 | 0.625 | -0.556 | -0.278 | 0.385 | -0.500 | 0.500 | -0.455 | -0.294 | 0.333 |
| 75 | S50 | -0.435 | 1.667 | — | 0.811 | 0.968 | 0.000 | 1.875 | — | 0.833 | 1.154 | -0.500 | 2.000 | — | 1.176 | 1.667 |
| 76 | G・SS | 0.435 | — | — | 0.270 | -0.323 | 0.800 | — | — | 0.278 | -0.385 | 0.000 | — | — | 0.294 | 0.667 |
| 77 | FT120 | 0.870 | 0.000 | -0.455 | 0.811 | -0.323 | 1.600 | 0.000 | -0.556 | 0.556 | 0.000 | 2.000 | -0.500 | 0.455 | 0.588 | -0.333 |
| 78* | FT120 | 0.435 | 0.556 | — | 1.351 | 0.968 | 0.400 | 0.625 | — | 0.833 | 0.769 | 1.000 | -0.500 | — | 0.294 | -0.333 |
| 79 | FT120 | -0.435 | — | — | -1.351 | -1.613 | -0.800 | — | — | -1.111 | -0.769 | 0.500 | — | — | -0.882 | -1.333 |
| 80 | FT120 | 1.304 | 0.556 | -0.455 | 1.351 | 0.323 | 1.600 | 1.875 | -0.556 | 0.833 | -0.385 | 1.500 | 1.000 | 0.000 | 0.294 | -0.667 |
| 81 | FT120 | -0.435 | -0.556 | 0.909 | 0.541 | 0.645 | -0.400 | 0.000 | 1.111 | 0.278 | 0.385 | 0.000 | 0.000 | 0.909 | -0.294 | -0.333 |
| 82 | S50 | 0.000 | 2.222 | — | 1.081 | 0.968 | 0.400 | 2.500 | — | 0.833 | 1.154 | 0.500 | 2.000 | — | 1.176 | 1.000 |
| 83 | FT2 | 0.000 | 0.000 | 0.909 | 1.081 | 1.290 | 0.400 | 0.000 | 0.000 | 0.833 | 1.538 | 0.500 | 0.000 | 0.000 | 1.176 | 1.667 |
| 84 | S50 | -0.870 | 0.556 | 1.818 | -0.811 | -0.968 | -0.400 | 0.625 | 2.222 | -0.833 | -0.769 | 0.500 | 0.000 | 1.818 | 0.000 | -0.333 |
| 85 | FT120 | -0.435 | 0.000 | 0.455 | 0.000 | 0.000 | 0.000 | 0.625 | 1.111 | 0.278 | 0.385 | -1.000 | 1.000 | 2.273 | 0.294 | 1.000 |
| 86 | ラクトスコープFTIR | 0.435 | 1.667 | — | -0.541 | -1.290 | 1.200 | 1.875 | — | 0.000 | -1.154 | 1.500 | 2.000 | — | 1.176 | 0.333 |
| 87 | S50 | 0.435 | -1.111 | 0.909 | 1.081 | 0.645 | 0.800 | 1.875 | -0.556 | 1.389 | 1.154 | 0.500 | 2.500 | 0.455 | 2.059 | 2.000 |
| 88 | S50 | 0.000 | 2.222 | -0.455 | 0.541 | 0.323 | 0.400 | 0.000 | -0.556 | 0.833 | 1.154 | 0.500 | 0.000 | -0.455 | 1.176 | 1.000 |
| 89 | FT120 | -0.435 | 0.000 | -0.455 | -0.270 | 0.000 | -0.400 | 0.000 | -0.556 | -0.556 | -0.385 | 0.500 | -1.500 | -0.909 | -0.882 | -1.333 |
| 90 | FT120 | -0.435 | 0.000 | 0.909 | 0.270 | 0.323 | -0.400 | 0.625 | 0.556 | -0.278 | 0.000 | 0.000 | 0.000 | 0.455 | -0.882 | -1.000 |
| 91 | ラクトスコープ | -0.870 | -0.556 | -1.818 | -0.811 | -0.645 | -0.400 | -0.625 | -1.111 | -0.556 | -0.385 | 1.500 | 0.500 | -0.909 | 1.176 | 0.333 |
| 92 | FT120 | 1.304 | 0.556 | 0.455 | 1.351 | 0.323 | 1.600 | 1.250 | 1.111 | 1.389 | 0.769 | 2.000 | 0.000 | 0.455 | 0.294 | -0.667 |
| 93 | G・SS | 0.435 | — | — | 1.622 | 1.290 | 2.000 | — | — | 0.278 | -1.538 | -1.000 | — | — | -0.882 | -0.333 |
| 94 | S50 | -0.435 | 1.667 | -0.909 | 0.541 | 0.645 | -0.800 | 1.250 | -0.556 | 0.000 | 0.769 | -0.500 | 1.000 | -0.455 | 0.294 | 0.667 |
| 95 | S50 | -1.304 | -0.556 | — | -1.081 | -0.323 | -0.800 | -1.250 | — | -1.111 | 0.385 | -1.500 | -1.000 | — | -0.882 | -0.667 |
| 96 | FT2 | 0.000 | -1.667 | 0.909 | -2.973 | ※ | 0.000 | -2.500 | 1.111 | -3.056 | ※ | 1.000 | -2.000 | 0.455 | -2.647 | -3.000 |
| 98 | G・TMSC | ※ | — | — | 0.541 | ※ | -2.400 | — | — | 1.944 | ※ | -1.500 | — | — | 0.000 | 1.000 |
| 102 | FT120 | 2.609 | 0.556 | 0.455 | 2.162 | 0.323 | 2.000 | 0.000 | -0.556 | 0.556 | -1.154 | 0.000 | 0.000 | 0.000 | -0.882 | -1.000 |

記号説明

- G ゲルベル法
- SS スマートシステム
- TMSC TMSチェッカー

注) ※は検定にて除外された値のため対象外

注) 番号に*がついた施設は期日までに報告がなかったため参考データとなります